

SPECIFICATION

JUNSEI CHEMICAL CO.,LTD.
QUALITY ASSURANCE SECTION
1-6, Ohmano-cho, Koshigaya-shi,
Saitama, 343-0844, Japan
TEL: +81-48-986-6161
FAX: +81-48-989-2787

Product Code 60176-4530
Commodity pH4.01 Standard buffer solution
Grade
Q'ty 500ml

[Test Item]	[Standard]
pH (25)	4.008 ± 0.005